

DECLARATION OF CONFORMITY EC

No. 66/D/EN/3

MANUFACTURER

REH4MAT

PRODUCT

ICE AND COMPRESSION THERAPY WRAP

Model: TB-01; TB-02; TB-03; TB-04; TB-05; TB-06; TB-07; TB-08; TB-09; TB-10;
TB 11; TB-12; TB-13; TB-14; TB-15; TB-16; TB-17; TB-18; TB-19; TB-20

marked with the CE sign is Class I medical device, of rule 1, consistent with the requirements of the **Council Directive 93/42/EEC of 14 June 1993 on medical devices** and the **Law of 20 May 2010 on medical devices** as amended. Conformity assessment was carried out on the basis of the **VII Council Directive 93/42/EEC**.

The product meets the requirements of the harmonized standards

PN-EN 1041+A1:2013-12

Information supplied by the manufacturer of medical devices

PN-EN 12182:2012

Assistive products for persons with disability. General requirements and test methods

PN-EN ISO 13485:2016

Medical Devices - Quality Management Systems - Requirements for Regulatory Purposes

PN-EN ISO 14971:2012

Medical devices. Application of risk management to medical device

PN - EN ISO 15223-1:2017-02

Medical devices Symbols to be used with medical device labels, labelling and information to be supplied - Part 1:General requirements

THIS DECLARATION HAS BEEN ISSUED ON THE RESPONSIBILITY
OF THE MANUFACTURER

REH4MAT, 36-060 Głogów Małopolski, ul. Piaski 47, Poland



rehamat 
COMPANY OWNER
Sławomir Wronski